

6.1 Specifications & Scope of Work (Revised 02/28/19)

DENTAL HYGIENE EQUIPMENT LIST (Revised 02/28/19)

Manufacturer	Model #	DESCRIPTION	Qty
DENTAL HYGIENE CLINIC			
Midmark	153758-001	MIDMARK ULTRATRIM CHAIR (ULTRALEATHER- Admiral Part Number 401)	18
Midmark	002-0837-00	CONTOURED JUNCTION BOX, HOUSING ONLY	18
Midmark	153964	LED TRACK LIGHT AND MONITOR (24" DENTAL GRADE MONITOR AND REMOTE CONTROL)	18
Midmark	CUSTOM	SYNTHESIS CASEWORK for 18 operatories - sized for each operatory w/specific arrangement 18 gauge steel frame, soft-close, full-extension, ball bearing glides, soft close concealed, 110 degree opening hinges, seamless panels & polystyrene drawers	18
Midmark	153659-002	MIDMARK FSTC WORK SURFACE for each operatory (12 O'clock delivery) ASSISTANT'S INSTRUMENTS, AND PROCENTER DELIVERY SYSTEM. INCLUDES 3 HANDPIECE POSITIONS WITH AIR/WATER SYRINGE ON DOCTORS SIDE AND SINGLE HVE, SALIVA EJECTOR, AND AIR/WATER SYRINGE ON ASSISTANTS GROUP. WORKSURFACE COMES IN CORIAN SOLID SURFACE. ALSO INCLUDES SELF CONTAINED WATER SYSTEM, AIR & WATER QUICK CONNECTS, AND WET/DRY FOOT CONTROL. Additional HVE on Dr. op	18
Midmark	CUSTOM	SYNTHESIS T WALL SINKS	8
Midmark	153811	MIDMARK DENTIST'S STOOL CONTOURED SEAT WITH RAISED FRONT/ WATERFALL FRONT EDGE/ ULTRALEATHER	18
Midmark		ASSISTANT'S STOOL (Color Admiral)	6
Polaroid		POLAROID 1001-10W WIRELESS INTRAORAL CAMERA	6
DENTAL HYGIENE MECHANICAL ROOM			
Midmark	G7	POWERVAC G7 DRY VACUUM 7-10 USER	2
Midmark	P72	MIDMARK P72 PWRAIR OILLESS 4.8HP CMPR MODEL: P72, NUMBER OF USERS: 7-10, NUMBER OF HEADS: 3, CFM@80PSI: 15.9 TANK (GAL.): 32, TOTAL HP: 4.8, VOLTAGE: 208-230, 60 HZ, TOTAL AMPS: 29.0, DBA: 73 DBA WITH SOUND COVER: 68, OVERALL DIMENSIONS H X W X D: 33" X 40" X 27", SOUND COVER OVERALL DIMENSIONS H X W X D: 39" X 40" X 30"	2

Midmark	30186000	With 1 TANDEM COMPRESSOR KIT	1
Midmark	77001598	And 1 MIDMARK CONTROL PANEL/CONTOL PANEL LOW VOLT AIR/AIR/VAC/VAC/WATER	1
Midmark	77002149	WATER FILTER/BY-PASS SYSTEM & Installation SOLENOID AND LOW VOLTAGE CONTACTOR, 1-1/2" PIPING, MAXIMUM GPM 40	1
Midmark	CUSTOM	SYNTHESIS SET-UP ROOM	1
Midmark	CUSTOM CUSTOM	STERILIZATION ROOM SYNTHESIS STERILIZATION CENTER PASS THROUGH INSTRUMENT CASSETTE CUBBIES	1
Midmark	M11-020	M11 ULTRACLAVE STERILIZER STERILIZER WITH AUTOMATIC DOOR. CHAMBER SIZE: 11" X 18" - PRE-PROGRAMMED CONTROLS: FOUR PRE-PROGRAMMED STERILIZATION CYCLE CONTROLS: UNWRAPPED (132° C OR 270° F FOR 3 MINUTES), POUCHES (132° C OR 270° F FOR 5 MINUTES), PACKS (121° C OR 250° F FOR 30 MINUTES), HANDPIECES (UNWRAPPED AND POUCHED) (132° C OR 270° F FOR 6 MINUTES).	3
Pelton & Crane	2969 65594 51646	VESTA PURE WATER FILTRATION SYSTEM & Installation VISTA PURE WATER STORAGE TANK 4.25 GALLONS NET VISTAPURE ELBOW TANK SHUT-OFF VALVE AUTOCLAVE WAND BOTTLE FAUCET	1 2 2 1 1
ESMA	E789	ESMA ULTRASONIC CLEANER & INSTALLATION IN COUNTER/CABINET TANK: 13"X11"X11" 6.5 GALLON - HOLDS 6 HU-FREIDY 8X11 CASSETTES	3
Midmark	CUSTOM	DENTAL MATERIALS MINI LAB SYNTHESIS LAB SET UP	1
Handler	31	#31 1/3 HP MODEL TRIMMER W/ PROTECTIVE SHIELD	2
Handler	26A	26A RED WING DENTAL LATHE	1
Handler	550-E	HANDLER 550 PORTA-VAC BENCH TOP DUST COLLECTOR	1
Keystone	7000352	THE VIBRATOR 110V	4
Buffalo	80186	ECONO-VAC VACUUM FORMER-120V AC, HIGH POWERED	3
		PLASTER TRAPS	3
		DENTAL RADIOLOGY AREA	

MIDMARK		PROGENY VANTAGE DIGITAL PANORAMIC SYSTEM	1	
MIDMARK		PREVA INTRAORAL DC XRAY SYSTEM	4	
Midmark		PHOSPHOR PLATE IMAGING KITS- Includes		
	CR-FP-12-012	Size 0 Imaging Plate Kit	2	
	CR-FP-12-013	Size 1 Imaging Plate Kit	4	
	CR-FP-12-014	Size 2 Imaging Plate Kit	6	
	CR-FP-12-020	Hygienic Bags Size 0 (Box of 100)	2	
	CR-FP-12-021	Hygienic Bags Size 1 (Box of 100)	2	
	CR-FP-12-022	Hygienic Bags Size 2 (Box of 100)	6	
Midmark	XPDZ-01	CLEARVISION CR SYSTEM PROCESSES SIZE 0,1,2,3, AND SIZE 4 PLATES. INCLUDES 4 SIZE 2 PLATES & BARRIERS AND PROGENY SOFTWARE.	3	Added
ALTERNATE #1:				
Midmark	XPDZ-01	CLEARVISION CR SYSTEM PROCESSES SIZE 0,1,2,3, AND SIZE 4 PLATES. INCLUDES 4 SIZE 2 PLATES & BARRIERS AND PROGENY SOFTWARE.	1	Revised Qty

Additional Requirements:

- NO SUBSTITUTIONS
- The Clearvision CR System XPDZ-01 shall be bid as Alternate #1
- Equipment storage until installation is the vendor's responsibility
- All delivery, freight and installation costs included, but list separately at the end of the bid
- Vendor responsible for debris removal
- Full (2) year warranty on parts & labor
- No charge quarterly preventative maintenance checks for first (2) years
- Emergency service response time between the hours of Monday-Friday, 7:00 am – 3:00 pm shall be (2) hours or less
- See Casework Drawings #1 & #2 & HS&HS Drawings